



MODEL AUDCAL

AUDIOMETER CALIBRATION SYSTEM

TYPICAL APPLICATIONS

- Audiometer calibration
- Booth qualification for ambient noise
- Speaker testing

PORTABLE, APP-CONTROLLED CALIBRATION

The SoundAdvisor™ Audiometer Calibration System AudCal is a field-ready solution for fast, comprehensive audiometer calibration. The system combines the power and versatility of the SoundAdvisor Model 831C sound level meter with the accessibility and functionality of the LD G4 Utility Software. AudCal works seamlessly with a tablet for both testing and data collection—no PC is required. Artificial ear couplers and mastoids are available to simplify the challenge of interfacing with a wide variety of audiometer transducers. Whether you are performing a typical audiometric test, a booth qualification for ambient noise, or a speaker test, Larson Davis is here to simplify your measurement challenges.

AudCal creates a complete, portable audiometer test and calibration solution. Access the meter on a PC or through the app on your mobile device to setup and run the full suite of tests with ease. Measurement capabilities include frequency, hearing level, FM, and pulse. Narrowband FFT and real time 1/3 octave analysis make measurements such as total harmonic distortion (THD) and ambient noise easy and accurate.

After completing the test suite, generate a test report and certificate that details the pass/fail, and parameters of each test. Test reports can be exported as a PDF that can be printed, shared, and saved.

SIMPLIFYING YOUR TESTING

When you need to test and measure an audiometer's performance to ensure accurate test results, Larson Davis is ready to support you with a fully inclusive Audiometer Calibration System. AudCal Systems deliver a complete package of hardware, firmware, and software designed to streamline your testing.

Packaged in a rugged case as a turnkey solution, AudCal Systems eliminate setup time by allowing audiometer tests to be performed quickly and easily right in the box. Whether you are performing a typical audiometric test, a booth qualification for ambient noise, or a speaker test, Larson Davis is here to simplify your measurement challenges.




MULTI-FUNCTIONAL SYSTEM

- **For All Your Transducers** – test virtually every device including supra-aural & circumaural headsets, bone vibrators, insert earphones, speakers, and sound booths
- **Couplers to Simplify** – artificial ear couplers and accessories including AEC201-A “318” ear simulator, AEC100 “NBS 9/A” coupler, AMC493C artificial mastoid, and AEC304 occluded ear simulator
- **Qualify Booths** – audiometric booth ambient noise with real-time 1/3 octave analysis and fast pass/fail results



USER FRIENDLY SOFTWARE

- **Intuitive Software** – AudCal Software leads you through each step with automatic detection of signals using four different standards including ANSI S3.6 (2010) and IEC 60645 (2017)
- **Store and Recall Tests** – System offers comprehensive search capabilities of previous tests
- **Reports are Easy** – print, email, share, or store reports and certificates for your client or your archives



ACCESSIBLE APP

- **Mobile Connectivity** – setup, control, and run your tests using the AudCal app
- **iOS and Android** – download the latest version of the app and connect your device to 802.11 WiFi to take control
- **Control the Test Environment** - See live data, add/edit customers, progress through the full test suite, and generate reports/certificates



ACOUSTIC COUPLERS & EAR SIMULATORS



SUPRA-AURAL EARPHONES

NBS 9 A 6CC COUPLER MODEL AEC100

- Designed for 1-inch microphones (not included) and the calibration testing of supra-aural earphones
- Delivers accurate and repeatable measurements up to 8 kHz
- Used for production testing where correlation between the coupler and real ear response is not a requisite



SUPRA-AURAL & CIRCUMAURAL EARPHONES

EAR SIMULATOR MODEL AEC201-A

- Used at frequencies up to 16000 Hz
- Meets IEC 60318-1:2009 Edition 2 and ANSI S3.7 section 5.4 requirements
- Compatible with earphones like TDH 39, TDH 49, TDH 50, HDA200 and Koss HV/1A
- Supplied with 377B13 microphone and Type 1 adapter plate (Type 2 available)
- Packaged with weights and accessories in a weather-tight case



INSERT HEARING AIDS AND EARPHONES

COUPLER MODELS AEC202 & AEC203


- Use AEC202 for ½ inch microphones
- Used for 1 inch microphones
- Both units meet IEC 60126 and IEC 60318-5 requirements
- AEC202 meets ANSI S3.7 2cc and AEC203 meets ANSI S3.7



INSERT EARPHONES

OCLCLUDED EAR SIMULATOR MODEL AEC304

- Designed to test insert earphones
- Includes ½ inch 12.5 mV/Pa matched microphone
- Meets IEC 60318-4 and IEC 60711:1981 requirements
- Often referred to as a Series 711 Coupler



BONE VIBRATOR TESTS

ARTIFICIAL MASTOID MODEL AMC493C

The AMC493C artificial mastoid is a precision mechanical coupler used to calibrate bone conduction hearing aids and audiometer bone vibrators. The AMC493C is cost effective and simple to use. Its patented design converts the vibrator force output to an acoustic signal measured with the system's sound level meter. It is used with the AEC100 coupler or AEC201-A Ear Simulator to perform bone vibrator tests.

| WHICH COUPLER SHOULD I USE FOR CALIBRATING AUDIOMETERS? | | | | | | | |
|---|--------|----------|--------|--------|-----------------|--------------------------------------|--|
| Head phone | AEC100 | AEC201-A | AEC202 | AEC203 | AEC304 | RETSPL | Notes |
| Ear Tone ER-3A/5A | | | ✓ | ✓ | ✓ (Occluded) | ISO 389-2 ANSI S3.6 | |
| Koss HV/1A | | ✓ | | | | ISO 389-5 ANSI S3.6 | Use 9-10 N weight and optional AEC201-2 |
| Telephonics TDH-39 3 | ✓ | ✓ | | | | ISO 389-1 ANSI S3.6 | Use 4-5 N weight |
| Telephonics TDH-49 | ✓ | ✓ | | | | ISO 389-1 ANSI S3.6 | Use 4-5 N weight |
| Telephonics TDH-50 | ✓ | ✓ | | | | ISO 389-1 ANSI S3.6 | Use 4-5 N weight |
| Sennheiser HDA200 | | ✓ | | | | ISO 389-5 ISO 389-8 ANSI S3.6 | Use 9-10 N weight and type 1 adapter plate |
| Sennheiser HDA280 | ✓ | ✓ | | | | ISO 389-1 ANSI S3.6 Sennheiser | Use 4-5 N weight |
| Sennheiser HDA300 | ✓ | ✓ | | | | Sennheiser | Use 4-5 N weight |
| Beyer DT-48 | ✓ | ✓ | | | | ISO 389-1 ANSI S3.6 | Use 4-5 N weight |
| Interacoustics DD45 | ✓ | ✓ | | | | Interacoustics | Use 4-5 N weight |
| Radio Ear B-71 | ✓ | ✓ | | | | ISO 389-3 ANSI S3.6 | Use 4-5 N weight and optional AMC493C weighting mass |

| MODEL NUMBER | DESCRIPTION |
|--|--|
| System Components | |
| SoundAdvisor 831C with 831C-AUD | Integrating Precision SLM (Type 1) with Low Noise Preamplifier (PRM831), NiMH Battery, Firmware for Testing Linearity, Frequency Response, FM, Pulse, Booth Noise, etc. |
| CAL250 | Class 1 acoustic calibrator, 114 dB, 250 Hz, 1 in opening, ½ in adapter (ADP019) |
| AEC100 | Artificial ear coupler (6cc) for 1 in microphone (NBS 9A) with adapter, weight, pillow |
| AEC201-A | Ear simulator with 377C13 microphone, adapter, weight, pillow |
| AEC202 | Artificial ear coupler (2cc) for ½ in microphone |
| AEC203 | Artificial ear coupler (2cc) for 1 in microphone |
| AEC304 | IEC 60711:1981 ear simulator including ½ in microphone |
| AMC493C | Artificial mastoid including storage humidior |
| Standards | |
| SoundAdvisor 831C | ANSI S1.4, IEC 61672-1, IEC 60651 and 60804 Class 1, ANSI S1.11 and IEC 61260 Class 1 |
| 831C-AUD | Measures requirements from ANSI S3.6 and IEC 60645 specifications for audiometers, ANSI S3.7 methods for coupler calibration of earphones, ANSI S3.1 maximum permissible ambient noise level for audiometer test rooms |

| MODEL NUMBER | DESCRIPTION |
|------------------------------|---|
| Standards (Continued) | |
| AEC100 | ANSI S3.7-1995, IEC 60318-3:1998 |
| AEC201 | ANSI S3.7 section 5.4, IEC 60318-1: 2009, directive 2004/108/EC |
| AEC202 | IEC 60318-5:2006 |
| AEC203 | IEC 60318-5:2006 |
| AEC304 | IEC 60318-4:2010 |
| AMC493C | ANSI S3.13-1981(R2007), IEC 60318-6: 2007, note: patented low thermal mass design varies from design features in standard |
| Physical | |
| SYS0xx | Weight: 22 lb (10 kg), CCS055 dimensions: 20 ½ x 16 ¾ x 8 ½ in (520 x 425 x 216 mm) |
| AEC201-A | Weight: 3.2 lb (1.4 kg) |
| AMC493C | Weight: 0.2 lb (0.05kg) |

| AUDCAL SYSTEM CONFIGURATIONS | | | | | | |
|-------------------------------------|--|----------|----------|----------|----------|----------|
| Item | Description | SYS014 | SYS015 | SYS016 | SYS017 | SYS-UPG |
| SLM | Precision sound level meter | 831C | 831C | 831C | 831C | 831C |
| Firmware | Audiometer calibration firmware | 831C-AUD | 831C-AUD | 831C-AUD | 831C-AUD | 831C-AUD |
| Calibrator | Class 1 acoustic calibrator | CAL250 | CAL250 | CAL250 | CAL250 | |
| Cable | 6 ft (2 m) extension cable | EXC006 | EXC006 | EXC006 | EXC006 | EXC006 |
| Case | Custom carrying case | CCS055 | CCS055 | CCS055 | CCS055 | CCS055 |
| DVX014 | WiFi adapter for 831C | ✓ | ✓ | ✓ | ✓ | ✓ |
| ADP010 | For measuring electrical and ambient noise | ✓ | ✓ | ✓ | ✓ | |
| AEC100 | NBS 9-A coupler for 1 in microphone | ✓ | ✓ | | | |
| 377A15 | 1 in pre-polarized pressure microphone | ✓ | ✓ | ✓ | ✓ | ✓ |
| ADP008A | Adapter to mount 1 in microphone onto PRM831 preamplifier | ✓ | ✓ | ✓ | ✓ | ✓ |
| AEC201-A | IEC 60318-1:2009 ear simulator with 377C13 ½ in microphone | | | ✓ | ✓ | |
| AMC493C | Artificial mastoid for bone vibrator test | | ✓ | | ✓ | |



3425 Walden Avenue, Depew, NY 14043 USA

larsondavis.com | sales@larsondavis.com | 888 258 3222 | +1 716 926 8243

© 2021 PCB Piezotronics - all rights reserved. PCB Piezotronics is a wholly-owned subsidiary of Amphenol Corporation. Endevco is an assumed name of PCB Piezotronics of North Carolina, Inc., which is a wholly-owned subsidiary of PCB Piezotronics, Inc. Accumetrics, Inc. and The Modal Shop, Inc. are wholly-owned subsidiaries of PCB Piezotronics, Inc. IMI Sensors and Larson Davis are Divisions of PCB Piezotronics, Inc. Except for any third party marks for which attribution is provided herein, the company names and product names used in this document may be the registered trademarks or unregistered trademarks of PCB Piezotronics, Inc., PCB Piezotronics of North Carolina, Inc. (d/b/a Endevco), The Modal Shop, Inc. or Accumetrics, Inc. Detailed trademark ownership information is available at www.pcb.com/trademarkownership.

MD-0340 revB 0222